

## 1. PERFORMANCE


- |                          |   |           |
|--------------------------|---|-----------|
| 1) Measuring range       | : 2-20 ppm  | 1-10 ppm  |
| Number of pump strokes   | 1/2 (50mℓ)  | 1 (100mℓ) |
| 2) Sampling time         | : 1 minute/1 pump stroke                                    |           |
| 3) Detectable limit      | : 0.2 ppm (100mℓ)   |           |
| 4) Shelf life            | : 3 years   |           |
| 5) Operating temperature | : 0 ~ 40 °C   |           |
| 6) Reading               | : Direct reading from the scale calibrated by 1 pump stroke |           |
| 7) Colour change         | : Pale purple → Pale yellow                                 |           |

## 2. RELATIVE STANDARD DEVIATION

RSD-low : 10 %    RSD-mid. : 5 %    RSD-high : 5 %

## 3. CHEMICAL REACTION

By reacting with Phosphoric acid, PH indicator is discoloured.



## 4. CALIBRATION OF THE TUBE

DIFFUSION TUBE METHOD

## 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	Coexistence
Ammonia	Similar stains are produced and higher readings are given.	Higher readings are given.
Other amines	∕	∕

(NOTE)

In case of 1/2 pump strokes, following formula is available for the actual concentration.

Actual concentration = 2 × Reading value